

PACE Model 203

DUAL CHAMBER (DDD)
TEMPORARY CARDIAC PACEMAKER



Now available with
terminals for bipolar
connectors

PACE Model 203 Highlights

- Constant voltage output for reliable cardiac capture (similar to permanent pacemakers)
- In Control: The first external pacemaker to instantly show the voltage (Volts) and current (mA) applied to the patient along with the impedance of the lead system (U.S. Patent 9,433,794):
- Check the quality and placement of the lead system (e.g., pacing wires);
- Ideally, the impedance of the lead system should be between 200 and 2,000 Ohms;
- Low impedance can be caused if stimulation electrodes are placed in close proximity or have a large surface area (loss of capture);
- High impedance may indicate a compromised connection or disconnected lead.
- P and R Wave amplitudes measured upon pressing PAUSE key
- Quick and easy change of battery: no need to worry about battery orientation (one single 9 V battery)
- Lifetime service



Specifications

PACE Model 203

Modes	DDD, VVI, AAI, VDD D00, V00, A00, DVI, DAI, VAT, AAT, DDD+AT, DAT	
Basic Pacing Rates	30 – 220 ppm	
Rapid Atrial Pacing Rates	70 – 1,000 ppm	
Output Amplitude	0.1 – 18 V	
Pulse Width	0.05 – 1.5 ms	
Sensitivity Threshold	Atrium : 0.2 – 20 mV	Ventricle: 1.0 – 20 mV
Refractory Period	250 ms	
Dimensions	P/N 2002: 21.7 cm x 9.6 cm x 5.1 cm (8.9" x 3.8" x 2.0")	P/N 2003: 21.2 cm x 9.6 cm x 5.1 cm (8.3" x 3.8" x 2.0")
Weight incl. Battery	490 g (17.3 ounces)	
Battery Type	9V Type 6LP3146 Alkaline	
Battery Life	Alkaline: Up to 10 days (72 ppm, 8.0V)	



Osyпка Medical GmbH

Albert-Einstein-Strasse 3
12489 Berlin, Germany
Phone: +49 (30) 6392 8300
Fax: +49 (30) 6392 8301
E-Mail: mail@osypkamed.com
www.osypkamed.com

United States of America:
Osyпка Medical, Inc.

dba. Cardiotronic
La Jolla, CA 92037, USA
Phone: +1 (858) 454 0600
Fax: +1 (858) 454 0640
E-Mail: mail@cardiotronic.net
www.cardiotronic.net

Copyright © 2017 Osypka Medical

